

Nano Science, Technology and Industry Scoreboard

Detecting SARS-COV-2 in 30 Seconds with Cheap and Sensitive Nanotubes

2020-09-16 Researchers from the University of Nevada, Reno demonstrate the power of cobalt-functionalized titanium dioxide nanotubes by selectively detecting the spike glycoprotein of the severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) in less than 30 seconds.

The relentless pandemic of coronavirus disease (COVID-19), caused by SARS-COV-2, resulted in a disruption to society not experienced since the 1918 influenza pandemic. One of the reasons for such rapid spread of the virus is the asymptomatic clinical presentation in many infected individuals, which contributes to involuntary transmission.

Consequently, consistent surveillance and quarantining of not only infected individuals but also asymptomatic ones could provide a somewhat effective measure to curtail further spread of SARS-CoV-2.

Nonetheless, methods that are currently being used for SARS-CoV-2 diagnosis are timeconsuming, expensive, and require trained personnel – albeit certain inexpensive solutions are obtaining approvals for emergency usage.

×

Schematic of Co functionalized TNT based sensing platform for the detection of SARS-CoV-2.

The promise of electrochemical biosensors

Electrochemical biosensors are a favorable choice for sensing biomolecules due to their propensity to detect biomarkers with adequate accuracy, high sensitivity, and specificity. As a result, they have been used to detect influenza viruses, enteroviruses, and Middle East

respiratory syndrome coronavirus (MERS-CoV).

The functionality of such electrochemical biosensors can be additionally improved by nanostructuring the electrode since this step increases the electrochemical reaction rate by increasing the electrode surface area to analyte fluid volume.

This is why researchers from the <u>University of Nevada</u> in Reno, hypothesized that SARS-CoV-2 could be detected through complexing of functionalized nanoparticles with characteristic viral proteins.

×

SEM micrographs of (a) TNTs post-annealing. Inset shows sidewalls of TNTs and (b) Co-functionalized TNTs showing the Co(OH)2 precipitate. Inset shows an EDS map of Co confirming its uniform distribution.

Synthesizing titanium dioxide nanotubes

This group of researchers, led by Dr. Bhaskar Sravan Vadlamani, developed a cobalt-metal functionalized titanium dioxide nanotubes (Co-TNT) as a sensing material for electrochemical detection of SARS-CoV-2 by detecting the receptor-binding domain (RBD) of spike glycoprotein.

More specifically, titanium dioxide nanotubes (TNTs) were synthesized by electrochemical anodization of titanium sheet. The annealed TNTs acquired from the furnace were functionalized with cobalt by employing an incipient wetting method, i.e., a wet ion exchange process.

The morphology of the TNTs and Co-TNT was then examined with the use of Dual Beam Scanning Electron Microscopy. Moreover, the spike-RBD protein, the aforementioned biomarker for SARS-CoV-2 detection, has been characterized via sodium dodecyl sulfatepolyacrylamide gel electrophoresis (SDS-PAGE) under denaturizing conditions.

The ability of Co-TNT to sense spike-RBD protein of SARS-CoV-2 was determined by carrying

out an amperometry experiment at -0.8 volts. Finally, the electrochemical sensing was performed with the use of a custom-built Co-TNT packaged and printed circuit board setup.

SARS-CoV-2 concentration and detection time

The study has shown that the sensor detected the spike-RBD protein of SARS-CoV-2 even at very low concentrations (ranging from 14 nM to 1400 nM). Furthermore, a linear response in the detection of viral protein with concentration has been demonstrated.

Importantly, the time needed for detection was very short (approximately 30 seconds), confirming the possibility of using this technology to develop a rapid diagnostic assay that can be exploited as a point of care diagnostics for rapid SARS-CoV-2 detection.

Also, the average sensor response time (defined as the time needed to reach the peak current) was found to be approximately 2 seconds. This is very short in comparison with earlier studies from the same authors on the sensor for colorectal cancer, as sensor response time there was 100-fold higher.

Scalable diagnostic platform

"Our report demonstrates the development of a simple, inexpensive, rapid, and non-invasive diagnostic platform that has the potential of detecting SARS-CoV-2 on clinical specimens including nasal, nasopharyngeal swabs or saliva" study authors summarize their findings.

"Moreover, the developed approach has the potential for diagnosis of other respiratory viral diseases by identifying appropriate metallic elements to functionalize TNTs," they add in this <u>medRxiv</u> preprint paper.

Advantageously, the sensor readout is an electrochemical signal gathered through a potentiostat or galvanostat, which can be adopted for smartphone applications and point of care diagnostics for COVID-19.

Finally, the sensor sensitivity may also be improved by using longer Co-TNT, since higher surface area provides a better reaction rate. As a result, higher sensor response current can be reached even at lower protein concentrations. All of this makes this technology a potentially useful addition to our diagnostic armamentarium.

Read the <u>original article</u> on Medical News.